

Keithley 2001 repair - Rework #283

No response from analog board

08/03/2012 02:18 PM - tin

Status: Closed	Start date: 08/03/2012
Priority: Normal	Due date:
Assignee: tin	% Done: 100%
Category:	Estimated time: 0.00 hour
Target version: Keithley 2001 #4	
Description Relays not working. Replaced U536 74HC00 fixed clock signal Replaced U309, U316, U310 comparators LM339 fixed heating MUX U325 issue	

History

#1 - 08/05/2012 01:47 PM - tin

- Status changed from *In Progress* to *Resolved*
- % Done changed from 90 to 100

Fixed.

U315 was replaced

#2 - 08/05/2012 02:38 PM - tin

- Status changed from *Resolved* to *Closed*