

Keithley 2001 repair - Feature #641

Finish AC input block schematics

12/04/2012 02:44 PM - tin

Status: Closed	Start date: 12/04/2012
Priority: Normal	Due date: 12/18/2012
Assignee: tin	% Done: 0%
Category:	Estimated time: 16.00 hours
Target version:	

Description
Finish recovery of mains power block on analog board.
Might be useful for unit #6 repair (might still have problem with 110VAC input, mains frequency detection circuitry)

History

#1 - 07/13/2014 01:32 AM - tin

- Status changed from New to In Progress
- Target version deleted (Keithley 2001 #6)

#2 - 02/24/2015 09:37 PM - tin

- Status changed from In Progress to Suspended

#3 - 04/07/2015 02:24 AM - tin

- Status changed from Suspended to Resolved

#4 - 04/11/2015 08:20 PM - tin

- Status changed from Resolved to Closed