

**Tektronix TDS/CSA7000 (CSA7404) - Bug #795**

**Channel 1 out of cal?**

09/08/2013 01:50 PM - tin

<b>Status:</b> Closed	<b>Start date:</b> 09/08/2013
<b>Priority:</b> Normal	<b>Due date:</b>
<b>Assignee:</b> tin	<b>% Done:</b> 100%
<b>Category:</b>	<b>Estimated time:</b> 0.00 hour
<b>Target version:</b>	
<b>Description</b>	

**History**

#1 - 09/09/2013 03:57 AM - tin

- Status changed from New to Resolved

- % Done changed from 0 to 100

Running SPC fixed errors, all channels are working same now.

Selftest reports 521 error 1.

#2 - 12/29/2013 12:21 AM - tin

- Status changed from Resolved to Closed